

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02086 Calib. Date: 10/13/2021 Calib. No.: 00044
Certification File No.: 02087 Cert. Date: 10/13/2021 Cert. No.: 00034
Linearity File No.: 02088 Lin. Date: 10/13/2021 Lin. No.: 00034
Solution File No.: 02096 Soln. Date: 12/17/2021 Soln. No.: 00432
Sequential File No.: 02096 File Date: 12/17/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 10/21/2022
Solution Control Lot: 20460 Bottle No.: 1488

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:46S		
Control 1 EC	0.100%	15:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:47S		
Control 2 EC	0.097%	15:48S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:48S		
Control 3 EC	0.098%	15:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: PTL. MCKAY #178

Badge No.: 178

Date: 12/17/2021