

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02021 **Calib. Date:** 10/13/2021 **Calib. No.:** 00038
Certification File No.: 02022 **Cert. Date:** 10/13/2021 **Cert. No.:** 00033
Linearity File No.: 02023 **Lin. Date:** 10/13/2021 **Lin. No.:** 00033
Solution File No.: 02044 **Soln. Date:** 12/17/2021 **Soln. No.:** 00377
Sequential File No.: 02044 **File Date:** 12/17/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 03/30/2022
Solution Control Lot: 20180 **Bottle No.:** 0628

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:41S		
Control 1 EC	0.101%	15:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:42S		
Control 2 EC	0.100%	15:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44S		
Control 3 EC	0.101%	15:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	15:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: _____

PTL MCKAY #178

Badge No.: 178

Date: 12/17/2021