

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 02086
Calib. Date: 10/13/2021
Calib. No.: 00044
Certification File No.: 02087
Cert. Date: 10/13/2021
Cert. No.: 00034
Linearity File No.: 02088
Lin. Date: 10/13/2021
Lin. No.: 00034
Solution File No.: 02098
Soln. Date: 12/31/2021
Soln. No.: 00433
Sequential File No.: 02098
File Date: 12/31/2021

Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDXB S3-113
Control Solution %: 0.100%
Expires: 10/21/2022
Solution Control Lot: 20460
Bottle No.: 1451

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:02S		
Control 1 EC	0.099%	16:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:03S		
Control 2 EC	0.096%	16:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:04S		
Control 3 EC	0.097%	16:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: PTL. MCKAY #178

Badge No.: 178

Date: 12/31/2021