

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02021 Calib. Date: 10/13/2021 Calib. No.: 00038
Certification File No.: 02022 Cert. Date: 10/13/2021 Cert. No.: 00033
Linearity File No.: 02023 Lin. Date: 10/13/2021 Lin. No.: 00033
Solution File No.: 02045 Soln. Date: 12/31/2021 Soln. No.: 00378
Sequential File No.: 02045 File Date: 12/31/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 03/30/2022
Solution Control Lot: 20180 Bottle No.: 0403

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:54S		
Control 1 EC	0.101%	15:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:55S		
Control 2 EC	0.099%	15:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:57S		
Control 3 EC	0.100%	15:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: PTL. MCKAY #178

Badge No.: 178

Date: 12/31/2021