

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02021 **Calib. Date:** 10/13/2021 **Calib. No.:** 00038
Certification File No.: 02022 **Cert. Date:** 10/13/2021 **Cert. No.:** 00033
Linearity File No.: 02023 **Lin. Date:** 10/13/2021 **Lin. No.:** 00033
Solution File No.: 02056 **Soln. Date:** 01/27/2022 **Soln. No.:** 00379
Sequential File No.: 02056 **File Date:** 01/27/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 03/30/2022
Solution Control Lot: 20180 **Bottle No.:** 0397

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:09S		
Control 1 EC	0.100%	05:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:10S		
Control 2 EC	0.100%	05:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:12S		
Control 3 EC	0.100%	05:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: PTL. MCKAY #178

Badge No.: 178

Date: 01/27/2022