

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 02086
Certification File No.: 02087
Linearity File No.: 02088
Solution File No.: 02107
Sequential File No.: 02107
Calib. Date: 10/13/2021
Cert. Date: 10/13/2021
Lin. Date: 10/13/2021
Soln. Date: 02/10/2022
File Date: 02/10/2022
Calib. No.: 00044
Cert. No.: 00034
Lin. No.: 00034
Soln. No.: 00437

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 20180
Model No.: CU-34
Serial No.: DDXB S3-113
Expires: 03/30/2022
Bottle No.: 0837

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:11S		
Control 1 EC	0.100%	10:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13S		
Control 2 EC	0.097%	10:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14S		
Control 3 EC	0.098%	10:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COX

First Name: ADAM

MI: C

Signature: _____

Badge No.: 164

Date: 02/10/2022