

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARZL-0146
Calibration File No.: 02021 Calib. Date: 10/13/2021 Calib. No.: 00038
Certification File No.: 02022 Cert. Date: 10/13/2021 Cert. No.: 00033
Linearity File No.: 02023 Lin. Date: 10/13/2021 Lin. No.: 00033
Solution File No.: 02060 Soln. Date: 02/10/2022 Soln. No.: 00380
Sequential File No.: 02060 File Date: 02/10/2022
Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 03/30/2022
Solution Control Lot: 20180 Bottle No.: 0515

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:02S		
Control 1 EC	0.100%	10:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04S		
Control 2 EC	0.099%	10:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05S		
Control 3 EC	0.100%	10:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COX

First Name: ADAM

MI: C

Signature: _____

Badge No.: 164

Date: 02/10/2022