

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARZL-0146
Calibration File No.: 02021
Calib. Date: 10/13/2021
Calib. No.: 00038
Certification File No.: 02022
Cert. Date: 10/13/2021
Cert. No.: 00033
Linearity File No.: 02023
Lin. Date: 10/13/2021
Lin. No.: 00033
Solution File No.: 02073
Soln. Date: 03/03/2022
Soln. No.: 00382
Sequential File No.: 02073
File Date: 03/03/2022

Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDXB S3-114
Control Solution %: 0.100%
Expires: 09/01/2023
Solution Control Lot: 21360
Bottle No.: 1327

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:41S		
Control 1 EC	0.100%	11:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42S		
Control 2 EC	0.100%	11:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44S		
Control 3 EC	0.098%	11:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COX

First Name: ADAM

MI: C

Signature: _____

Badge No.: 164

Date: 03/03/2022