

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	02081	Calib. Date:	03/08/2022
Certification File No.:	02082	Cert. Date:	03/08/2022
Linearity File No.:	02083	Lin. Date:	03/08/2022
Solution File No.:	02090	Soln. Date:	03/25/2022
Sequential File No.:	02090	File Date:	03/25/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	0904

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:32S		
Control 1 EC	0.099%	10:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:33S		
Control 2 EC	0.099%	10:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34S		
Control 3 EC	0.099%	10:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COX

First Name: ADAM

MI: C

Signature: _____

Badge No.: 164

Date: 03/25/2022