

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARZL-0146
Calibration File No.: 02081
Calib. Date: 03/08/2022
Calib. No.: 00039
Certification File No.: 02082
Cert. Date: 03/08/2022
Cert. No.: 00034
Linearity File No.: 02083
Lin. Date: 03/08/2022
Lin. No.: 00034
Solution File No.: 02099
Soln. Date: 04/15/2022
Soln. No.: 00385
Sequential File No.: 02099
File Date: 04/15/2022
Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDAB-0015
Control Solution %: 0.100%
Expires: 10/21/2022
Solution Control Lot: 20460
Bottle No.: 1480

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:52D		
Control 1 EC	0.099%	09:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54D		
Control 2 EC	0.098%	09:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:55D		
Control 3 EC	0.099%	09:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:57D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COX

First Name: ADAM

MI: C

Signature: 

Badge No.: 164

Date: 04/15/2022