

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02112 Calib. Date: 03/08/2022 Calib. No.: 00045
Certification File No.: 02113 Cert. Date: 03/08/2022 Cert. No.: 00035
Linearity File No.: 02114 Lin. Date: 03/08/2022 Lin. No.: 00035
Solution File No.: 02118 Soln. Date: 04/15/2022 Soln. No.: 00442
Sequential File No.: 02118 File Date: 04/15/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 10/21/2022
Solution Control Lot: 20460 Bottle No.: 1487

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:30D		
Control 1 EC	0.101%	09:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31D		
Control 2 EC	0.099%	09:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33D		
Control 3 EC	0.099%	09:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COX

First Name: ADAM

MI: C

Signature: _____

Badge No.: 164

Date: 04/15/2022