

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02081 Calib. Date: 03/08/2022 Calib. No.: 00039
Certification File No.: 02082 Cert. Date: 03/08/2022 Cert. No.: 00034
Linearity File No.: 02083 Lin. Date: 03/08/2022 Lin. No.: 00034
Solution File No.: 02108 Soln. Date: 05/13/2022 Soln. No.: 00386
Sequential File No.: 02108 File Date: 05/13/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 10/21/2022
Solution Control Lot: 20460 Bottle No.: 1439

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:19D		
Control 1 EC	0.100%	12:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21D		
Control 2 EC	0.099%	12:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22D		
Control 3 EC	0.099%	12:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COX

First Name: ADAM

MI: C

Signature: _____

Badge No.: 164

Date: 05/13/2022