

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02112 **Calib. Date:** 03/08/2022 **Calib. No.:** 00045
Certification File No.: 02113 **Cert. Date:** 03/08/2022 **Cert. No.:** 00035
Linearity File No.: 02114 **Lin. Date:** 03/08/2022 **Lin. No.:** 00035
Solution File No.: 02119 **Soln. Date:** 05/13/2022 **Soln. No.:** 00443
Sequential File No.: 02119 **File Date:** 05/13/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 10/21/2022
Solution Control Lot: 20460 **Bottle No.:** 1449

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:21D		
Control 1 EC	0.100%	12:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:23D		
Control 2 EC	0.098%	12:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24D		
Control 3 EC	0.099%	12:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COX

First Name: ADAM

MI: C

Signature: _____

Badge No.: 164

Date: 05/13/2022