

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 02081 **Calib. Date:** 03/08/2022 **Calib. No.:** 00039
Certification File No.: 02082 **Cert. Date:** 03/08/2022 **Cert. No.:** 00034
Linearity File No.: 02083 **Lin. Date:** 03/08/2022 **Lin. No.:** 00034
Solution File No.: 02114 **Soln. Date:** 06/01/2022 **Soln. No.:** 00387
Sequential File No.: 02114 **File Date:** 06/01/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0015
Control Solution %: 0.100% **Expires:** 09/01/2023
Solution Control Lot: 21360 **Bottle No.:** 1376

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:17D		
Control 1 EC	0.100%	14:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:18D		
Control 2 EC	0.099%	14:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:19D		
Control 3 EC	0.099%	14:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COX

First Name: ADAM

MI: C

Signature: _____

Badge No.: 164

Date: 06/01/2022