

# 148

**State of New Jersey  
Department of Law and Public Safety  
Division of Criminal Justice**

THIS IS TO CERTIFY THAT


**THOMAS C LASKO**

HAS SUCCESSFULLY COMPLETED THE PRESCRIBED PROGRAM AND IS HEREBY  
AUTHORIZED TO BE A

**Radar Operator/Instructor**

2/8/2012  
Issue Date

2/8/2015



For Division of Criminal Justice  
use only - discard this section

**Wall Township Police  
Monmouth**

**State of New Jersey  
Department of Law and Public Safety  
Division of Criminal Justice**

THIS IS TO CERTIFY THAT

**THOMAS C LASKO**

HAS SUCCESSFULLY COMPLETED THE PRESCRIBED PROGRAM  
AND IS HEREBY AUTHORIZED TO BE A

**Radar Operator/Instructor**

2/8/2012  
Issue Date

2/8/2015



# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00459	Calib. Date: 03/07/2012
Certification File No.:	00460	Calib. No.: 00010
Linearity File No.:	00461	Cert. Date: 03/07/2012
Solution File No.:	00469	Cert. No.: 00008
Sequential File No.:	00469	Lin. Date: 03/07/2012
		Lin. No.: 00008
		Soln. Date: 04/05/2012
		Soln. No.: 00086
		File Date: 04/05/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0508

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:19D		
Control 1 EC	0.103%	13:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20D		
Control 2 EC	0.100%	13:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:22D		
Control 3 EC	0.099%	13:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A.

Signature: \_\_\_\_\_



Badge No.: 129

Date: 04/05/2012

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00555	Calib. Date: 03/07/2012
Certification File No.:	00556	Calib. No.: 00015
Linearity File No.:	00557	Cert. Date: 03/07/2012
Solution File No.:	00570	Cert. No.: 00009
Sequential File No.:	00570	Lin. Date: 03/07/2012
		Lin. No.: 00009
		Soln. Date: 04/05/2012
		Soln. No.: 00112
		File Date: 04/05/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0505

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:13D		
Control 1 EC	0.101%	13:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14D		
Control 2 EC	0.101%	13:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15D		
Control 3 EC	0.100%	13:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:17D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A.

Signature: 

Badge No.: 129

Date: 04/05/2012